

Abstract

New HPLC Method for Surfactants Detection in Wastewaters Samples [†]

Iuliana Paun ^{1,2,*}, Laura Florentina Chiriac ¹, Florinela Pirvu ^{1,2}, Vasile Ion Iancu ¹ and Cristina Ileana Covaliu-Mierla ²

¹ National Research and Development Institute for Industrial Ecology – ECOIND, 57-73 Drumul Podu Dambovitei, Bucharest, 060652, Romania; laura.badea88@yahoo.com (L.F.C.); florinela_pirvu@yahoo.com (F.P.); vasileiancu10@gmail.com (V.I.I.)

² University POLITEHNICA of Bucharest, Faculty of Biotechnical Systems Engineering, 313 Splaiul Independentei, Bucharest, 060042, Romania; cristina_covaliu@yahoo.com

* Correspondence: iuliana_paunita@yahoo.com

† Presented at the 1st International Electronic Conference on Processes: Processes System Innovation, 17–31 May 2022; Available online: <https://ecp2022.sciforum.net>.

Abstract: Over the last decade, biocides have received increasing attention due to their widespread use, their transfer to aquatic ecosystems and their negative effect on aquatic organisms. Alkyl benzyl dimethyl ammonium chlorides are applied as bactericides and disinfectants in sanitary products and antistatic agents in the formula of laundry conditioners. The aim of this study was to provide a sensitive and robust HPLC-DAD method for detection of three biocides (dodecyl- (C12-), tetradecyl- (C14-), and hexadecyl- (C16-) benzyl dimethylammonium chloride) in wastewater samples. The analytes separation was achieved using an Acclaim Surfactant Plus (3 μ m, 150 mm \times 3 mm) chromatographic column, maintained at 30 °C. The isocratic mode elution using a binary phase of ammonium acetate 0,2 M (A): acetonitrile (B) as mobile phase (50:50, *v/v*) at a flow rate of 0.5 ml/min, allowed a run time of only 5 minutes. The linearity, accuracy and intermediate precision were validated. The HPLC-DAD method provides good linearity, with correlation coefficients from 0.9992 to 0.9997 in the concentration range from 1 to 100 mg/L. Very good precision values were obtained, with RSD% ranged from 1.37–2.27% for intra-day measurements and between 6.14 and 6.65% for inter-day measurements. The target biocides were isolated from wastewater samples through Solid Phase Extraction (SPE) procedure, using polymeric Strata-X Cartridges and acetonitrile and acetic acid (90%/10%) as elution solvent mixture. Recoveries (up to 86%) made possible the quantification biocides at very low levels, the limits of quantification (LOQs) being in the range of 4.5 and 7.6 μ g/L. The method was successfully applied to wastewater samples, obtaining concentration values varying from few μ g/L to few mg/L.

Citation: Paun, I.; Chiriac, L.F.; Pirvu, F.; Iancu, V.I.; Covaliu-Mierla, C.I. New HPLC Method for Surfactants Detection in Wastewaters Samples. *Proceedings* **2022**, *69*, x. <https://doi.org/10.3390/xxxxx>

Academic Editor: Firstname Lastname

Published: date

Publisher's Note: MDPI stays neutral with regard to jurisdictional claims in published maps and institutional affiliations.



Copyright: © 2022 by the authors. Submitted for possible open access publication under the terms and conditions of the Creative Commons Attribution (CC BY) license (<https://creativecommons.org/licenses/by/4.0/>).

Keywords: HPLC-DAD; biocides; solid phase extraction; wastewater samples