

The 1st International Electronic Conference on Chemical Sensors and Analytical Chemistry 01–15 JULY 2021 | ONLINE



## Eucalyptus biochar as a sustainable nanomaterial for electrochemical

sensors



<u>Annalisa Scroccarello<sup>1\*</sup></u>, Flavio Della Pelle<sup>1</sup>, Qurat Ul Ain Bukhari<sup>1</sup>, Filippo Silveri<sup>1</sup>, Daniele Zappi<sup>2</sup>, Enrico Cozzoni<sup>3</sup>, Dario Compagnone<sup>1</sup>



<sup>1</sup>Faculty of Bioscience and Technology for Food, Agriculture and Environment University of Teramo, Via Renato Balzarini 1, 64100, Teramo, Italy <sup>2</sup>Biosensor, Via Olmetti 44, 00060, Formello, Roma, Italy <sup>3</sup>GRADO ZERO INNOVATION, Via nove 2/A, 50056, Montelupo F.no, Firenze, Italy

\*ascroccarello@unite.it

## State of the art







Biochar: Production, properties and emerging role as a support for enzyme immobilization

Trends in Analytical Chemistry 128 (2020) 11590

Deepshikha Pandey, Achlesh Daverey<sup>\*</sup>, Kusum Arunachalam<sup>\*\*</sup>



UNIVERSITA

DEGLI STUD **DI TERAMO** 

Van der Waals force **Ionic Interaction** Enzyme Check for updates **Covalent** bonding **Cross Linker** 

Contents lists available at ScienceDirect Trends in Environmental Analytical Chemistry journal homepage: www.elsevier.com/locate/teac

Electrochemical devices obtained from biochar: Advances in renewable and environmentally-friendly technologies applied to analytical chemistry



Leandro S. de Almeida, Eliézer Q. Oreste, Juliana V. Maciel, Mônika G. Heinemann, Daiane Dias\*



Transa Ramanayaka, S., Vithanage, M., Alessi, D. S., Liu, W. J., Jayasundera, A. C., & Ok, Y. S. (2020). Environmental Science: Nano, 7(11), 3279-3302.

TrEAC

Xiao, X., Chen, B., Chen, Z., Zhu, L., & Schnoor, J. L. (2018). *Environmental science & technology*, 52(9), 5027-5047.

## Biochar preparation, electrode set-up and fabrication



### Preparation of Biochar dispersion through liquid-phase sonochemical approach



## **Biochar as sensing element: amount optimization**





CVs: 5 mM Fe(CN)<sub>6</sub><sup>4-/3-</sup> in 0.1M KCl; scan rate 25 mV s<sup>-1</sup>.

EIS: 5 mM Fe(CN)<sub>6</sub><sup>4-/3-</sup> in 0.1M KCI; frequency range of 10<sup>5</sup> - 10<sup>-1</sup> Hz using 5mV as sinusoidal wave amplitude, fixing the potential at open circuit.</sup>

## Biochar as sensing element: mono- and ortho-phenols evaluation

#### UNIVERSITÀ DEGLI STUDI DI TERAMO UNITE

### **Explorative cyclic voltammetries**



Cyclic voltammograms at 25 mV s-1 of phenol-containing molecule couples

## Biochar as sensing element: mono- and ortho-phenols evaluation

Analysis of Tyrosol and Hydroxytyrosol in extra virgin olive oil



UNIVERSITÀ

DEGLI STUD

DPV parameters: Pulse width 50 ms, modulation amplitude 50 mV, scan rate of 25 mV s<sup>-1</sup>. The samples were diluted 1:200 and 1:100 v/v in PB 0.1 M for BH-SPE and BH-film based electrode respectively.

## Conclusions





A water-soluble Biochar nanofiber dispersion was obtained



The Biochar nanofiber was used as SPE electrode modifier and to realize biochar-based conductive film

The conductive film was integrated into an entirely lab-made flexible electrode

The Biochar-based electrodes present great analytical performance for ortho-phenols compounds



The Biochar-based electrode allow the selective detection of ortho-phenols in presence of mono-phenols compounds



Sensors showed useful linear ranges and limits of detections

The sensors were applied in EVOO samples analysis and quantitative recoveries were obtained

## Acknowledgments



Prof. Dario Compagnone Dr. Flavio Della Pelle Ph.D. Filippo Silveri Ph.D. Qurat Ul Ain Bukhari



Dr. Maria Teresa Giardi Dr. Daniele Zappi



Dr. Enrico Cozzoni

# ... THANKs FOR YOUR ATTENTION!

UNIVERSITA DEGLI STUD

