

Comprehensive Optimization of the Tripolar Concentric Ring Electrode with Respect to the Accuracy of Laplacian Estimation Based on the Finite Dimensions Model of the Electrode [†]

Oleksandr Makeyev

School of STEM, Diné College, Tsaile, AZ 86556, USA; omakeyev@dinecollege.edu; Tel.: +1-928-724-6960

[†] Presented at the 7th International Electronic Conference on Sensors and Applications, 15–30 November 2020; Available online: <https://ecsa-7.sciforum.net/>.

Published: 15 November 2020

Abstract: Optimization performed in this study is based on the finite dimensions model of the concentric ring electrode as opposed to the negligible dimensions model widely used in the past. This makes the optimization problem comprehensive since all of the electrode parameters including, for the first time, the radius of the central disc and individual widths of concentric rings are optimized simultaneously. The optimization criterion used is maximizing the accuracy of the surface Laplacian estimation since ability to estimate Laplacian at each electrode constitutes primary biomedical significance of concentric ring electrodes. Even though obtained results and derived principles defining optimal electrode configurations are illustrated on tripolar (2 concentric rings) electrodes, they were also confirmed for quadripolar (3 rings) and pentapolar (4 rings) electrodes and are likely to continue to hold for any higher number of concentric rings. For tripolar concentric ring electrodes, the optimal configuration was compared to previously proposed linearly increasing inter-ring distances and constant inter-ring distances configurations of the same size and based on the same finite dimensions model of the electrode. Obtained results suggest that previously proposed configurations correspond to an almost two-fold and more than three-fold increases in Laplacian estimation error respectively compared to the optimal configuration proposed in this study.

Keywords: electrophysiology; measurement; wearable sensors; noninvasive; concentric ring electrodes; Laplacian; estimation; optimization

1. Introduction

Concentric ring electrodes (CREs) are noninvasive electrodes for electrophysiological measurement with primary biomedical significance tied to their ability to accurately estimate the Laplacian (second spatial derivative of the surface potential) at each electrode. Properties shared by the majority of currently used CREs are: relatively small radius of the central disc (compared to the radius of the electrode) and/or equal and small widths of concentric rings (compared to the radius of the electrode) [1–8]. These properties stem, at least partially, from the use of the negligible dimensions model (NDM) of a CRE – a Cartesian grid where the central disc is represented by a single point (of negligible diameter) in the middle of the grid and rings are represented by concentric circles (of negligible width) around it. For example, since NDM was used to calculate Laplacian estimates for tripolar CRE (TCRE) in [9], it also influenced the design of respective physical electrodes. Previous results on improving the Laplacian estimation accuracy via CRE optimization were also based on NDM [10–12].

First proof of concept of the finite dimensions model (FDM) of the CRE with nonnegligible widths of concentric rings and the radius of the central disc was introduced in [13] before being developed into a comparison framework validated on human electrocardiogram data in [5]. This framework, allowing direct comparison between two CRE configurations with the same number of rings and the same size but with different radii of the central disc, widths of concentric rings, and inter-ring distances, was used in this study to define and solve a comprehensive CRE optimization problem maximizing the Laplacian estimation accuracy via said CRE. Unlike NDM based optimization problem that was solved in [12], this study includes and optimizes all the CRE parameters simultaneously. Absolute values of truncation term coefficients of the lowest remaining order were compared since in [11] and [12] ratios of those coefficients have been shown, using finite element method modeling, to be predictors of the Laplacian estimation error. Specifically, ratios of Relative and Maximum Errors of Laplacian estimation calculated using finite element method modeling and analytic ratios of truncation term coefficients differed by less than 5% for combinations of linearly increasing inter-ring distances (LIIRD), constant inter-ring distances (CIRD), and linearly decreasing inter-ring distances TCRES and quadripolar CREs [11] as well as for their quadratically increasing inter-ring distances counterparts [12]. Moreover, in [5] consistency between NDM and FDM in terms of values of truncation term coefficient ratios have been demonstrated for CIRD and LIIRD TCRES configurations. This is to be expected since NDM and FDM are also consistent in terms of the highest order of the truncation term that can be cancelled out during derivation of the Laplacian estimate that has been shown to be equal to twice the number of concentric rings in the electrode in [10] (for NDM) and [13] (for FDM) respectively.

As a result of this study, general principles defining optimal CRE configurations maximizing the accuracy of Laplacian estimation are defined and illustrated for the case of TCRES. Moreover, optimal TCRES configuration is directly compared to LIIRD and CIRD configurations from [5]. CIRD configuration from [5] corresponds to a more than three-fold increase in Laplacian estimation error while LIIRD configuration from [5] corresponds to an almost two-fold increase in Laplacian estimation error compared to the optimal TCRES configuration proposed in this study.

2. Materials and Methods

2.1. Preliminaries

Figure 1 represents the FDM diagrams of three TCRES configurations including CIRD (Figure 1A) and LIIRD (Figure 1B) ones that were used to illustrate the comparison framework in [5]. All three configurations in Figure 1 have the same radius subdivided into 9 equal intervals. CIRD and LIIRD configurations have the radius of the central disc and widths of both rings equal to $1/9$ of the electrode radius. For CIRD configuration both distance between the central disc and the middle ring and distance between the middle ring and the outer ring are equal to $3/9 (= 1/3)$ of the electrode radius. For LIIRD configuration the distance between the central disc and the middle ring ($2/9$ of the electrode radius) is one half of the distance between the middle ring and the outer ring ($4/9$ of the electrode radius). Average potential on each concentric circle with the radius ranging from 1 to 9 is calculated using Huiskamp's Laplacian potential derivation based on the Taylor series expansion from [14]. Main steps of the comparison framework for TCRES configuration are listed below (see [5] for more detail) with similar steps used for quadripolar, pentapolar, etc configurations:

1. Calculating the potentials on all three recording surfaces (central disc and two concentric rings) of the TCRES. For example, the potential on the central disc in all three TCRES configurations in Figure 1 is equal to the average of the potential at the center of the central disc and the potential on the concentric circle with radius equal to $1/9$ of the electrode radius.
2. Canceling out the potential at the center of the central disc by taking bipolar differences between potentials on the middle ring and on the central disc and between potentials on the outer ring and on the central disc respectively.
3. Combining the two bipolar differences linearly to cancel out the 4th (twice the number of concentric rings) order truncation term, solve for the surface Laplacian estimate, and to calculate

the absolute value of the 6th order truncation term coefficient (lowest remaining truncation term order for TCRE). Lowest remaining truncation term order is used since “higher-order terms usually contribute negligibly to the final sum and can be justifiably discarded” from the Taylor series [15].




Figure 1. Finite dimensions models of three tripolar concentric ring electrode configurations including: (A) constant inter-ring distances configuration, (B) linearly increasing inter-ring distances configuration, and (C) optimal configuration with respect to the accuracy of Laplacian estimation.

2.2. Optimization Problem

Comparison framework from [5] has been developed into a comprehensive optimization problem directly comparing not pairs but all the possible CRE configurations of the same size and with the same number of rings simultaneously. Absolute values of truncation term coefficients for the lowest remaining truncation term order were calculated for CRE configurations of given electrode size and with given number of rings including all the possible combinations of values for the radius of the central disc, widths of concentric rings, and inter-ring distances. The lowest absolute value of the truncation term coefficient corresponds to the highest accuracy of Laplacian estimation and vice versa.

3. Results

3.1. General Principles Defining Optimal CRE Configurations

Before the general principles that define optimal CRE configurations maximizing the accuracy of Laplacian estimation are introduced, the results of optimization for TCRE with the outer radius of the outer ring (the electrode radius) equal to 6 are presented in Table 1. These results will be used to illustrate each of the aforementioned principles.

Table 1. All possible TCRE configurations for the outer radius of the outer ring equal to 6.

TCRE Number	Central Disc Radius	Middle Ring Radii		Outer Ring Radii		Absolute Value of the 6th Order Truncation Term Coefficient	Increase with Respect to the Optimal (%)
		Inner	Outer	Inner	Outer		
1	1	2	3	4	6	0.685	0
2	1	2	3	5	6	0.717	4.65
3	1	2	4	5	6	1.096	59.99
4	1	3	4	5	6	1.250	82.53
5	2	3	4	5	6	1.369	99.93

Table 1 contains all 5 possible TCRE configurations sorted in accordance with the respective absolute values of the 6th order truncation term coefficients whose ratios have been shown to be

predictors of the Laplacian estimation error in [11,12] (hence the two terms are used interchangeably below). Percentage of increase in the absolute value of the 6th order truncation term coefficient with respect to the optimal configuration (TCRE configuration number 1) is also provided in the rightmost column of Table 1. It can be seen from Table 1 that even for such small electrode radius of 6 (reducing it to 5 results in a single possible TCRE configuration) the difference between the Laplacian estimation errors for the optimal and the worst-case scenario TCRE configurations (TCRE configuration number 5) approaches 100%.

General principles defining optimal CRE configurations (independent of the number of concentric rings) in terms of accuracy of the surface Laplacian estimate are:

1. In the optimal configuration, central disc and concentric rings are kept at minimum distances with minimum radius/widths except for the width of the outer ring.
Example: TCRE configuration number 1 in Table 1.
2. Larger width of the outer ring is advantages to smaller width in electrode configurations that are otherwise identical.
Example: TCRE configuration number 1 versus number 2 in Table 1.
3. Increasing the width of a concentric ring closer to the outer edge of the electrode is advantageous to increasing the width of a concentric ring closer to the central disc.
Example: TCRE configuration number 2 versus numbers 1 and 3 in Table 1.
4. Increasing the width of any concentric ring is advantageous to increasing the radius of the central disc.
Example: TCRE configuration number 1 versus numbers 3 and 5 in Table 1.
5. Increasing the distance between recording surfaces closer to the outer edge is advantageous to increasing the distance between recording surfaces closer to the central disc.
Example: TCRE configuration number 2 versus number 4 in Table 1.

3.2. Comparison of the Optimal TCRE Configuration with Previous Results

Results of FDM optimization for TCRES are compared with our previously obtained results in Table 2 that corresponds to the outer radius of the outer ring (the electrode radius) equal to 9.

Table 1. Select TCRE configurations for the outer radius of the outer ring equal to 9.

TCRE Number	Central Disc Radius	Middle Ring Radii		Outer Ring Radii		Absolute Value of the 6th Order Truncation Term Coefficient	Increase with Respect to the Optimal (%)
		Inner	Outer	Inner	Outer		
1	1	2	3	4	9	1.447	0
2	1	2	3	5	9	1.458	0.78
3	1	2	3	6	9	1.489	2.94
4	1	2	3	7	9	1.550	7.19
5	1	2	3	8	9	1.650	14.07
...
15	1	3	4	8	9	2.883	99.33
...
30	1	4	5	8	9	4.528	213.01
...
66	4	5	7	8	9	9.189	535.22
67	2	6	7	8	9	9.407	550.35
68	3	6	7	8	9	9.901	584.45
69	4	6	7	8	9	10.436	621.46
70	5	6	7	8	9	10.879	652.05

Out of the total of 70 possible TCRE configurations with radius equal to 9 Table 2 presents the top 5, the bottom 5, and two TCRE configurations assessed in [5]: CIRD (TCRE configuration number 30; Figure 1A) and LIIRD (TCRE configuration number 15; Figure 1B). While the results in Table 2 follow the same general principles defining optimal CRE configurations as the results in Table 1, the

difference between the Laplacian estimation errors for the optimal and the worst-case scenario TCRE configurations increased to over 650% in Table 2 compared to under 100% in Table 1. This increase of more than 6.5 times is due to just 1.5 times increase in the electrode radius (from 6 in Table 1 to 9 in Table 2). More importantly, in direct comparison the optimal TCRE configuration (TCRE configuration number 1 in Table 2; Figure 1C) outperforms two TCRE configurations from [5] by 99.33% and 213.01% respectively in terms of the Laplacian estimation error.

4. Discussion

In this study optimizing the FDM based TCRE configurations with respect to the accuracy of Laplacian estimation (Tables 1 and 2), the distinctive feature of obtained results is that in optimal TCRE configurations the recording surfaces account for the vast majority of the electrode surface area minimizing the distances between the recording surfaces (e.g., optimal TCRE configuration in Figure 1C). This is markedly different from the currently used CREs where majority of the electrode surface area corresponds to the distances between the recording surfaces (for example, see CREs from [5,6]). Compared to the optimal TCRE configuration (Figure 1C), LIIRD configuration of the same size (Figure 1B) increases the Laplacian estimation error almost two-fold while CIRD configuration (Figure 1A) corresponds to a more than three-fold increase. These results have potential to inform the design of future CREs and could not have been obtained with simplistic NDM.

At the same time, some aspects of FDM based optimal configurations are consistent with the previous results obtained using NDM. For example, locating the middle ring closer to the central disc than to the outer ring is consistent with analytical and finite element method modeling results from [11,12]. Another example is an increase in Laplacian estimation error associated with an increase in the size of the electrode observed via modeling in [10–12]. Similar result can be seen in Tables 1 and 2 corresponding to the TCRE radius increasing from 6 to 9 with the optimal absolute value of the 6th order truncation term coefficient increasing from 0.685 to 1.447 respectively.

The only optimization criterion used in this study was maximizing the accuracy of surface Laplacian estimation via the CRE. Other optimization criteria may result in different optimal electrode configurations so adding additional optimization criteria to the optimization problem solved in this study is one of the directions of the future work. More importantly, for optimal CRE configurations the question of how small can the distances between the recording surfaces get before shorting due to salt bridges negatively affects the accuracy of Laplacian estimation becomes more critical than before since the first principle defining optimal configurations is to keep those distances minimal. Modeling or prototyping of optimal CRE configurations is needed to answer this question.

Funding: This research was funded by the National Science Foundation (NSF) Division of Human Resource Development (HRD) Tribal Colleges and Universities Program (TCUP), grant number 1914787 to O.M.

Acknowledgments: The author gratefully acknowledges Ernst Kussul and Tetyana Baydyk (National Autonomous University of Mexico, Mexico) for the constructive discussions and helpful comments.

Conflicts of Interest: The author declares no conflict of interest. The funders had no role in the design of the study; in the collection, analyses, or interpretation of data; in the writing of the manuscript, or in the decision to publish the results.

References

1. Wang, K.; Parekh, U.; Pailla, T.; Garudadri, H.; Gilja, V.; Ng, T.N. Stretchable Dry Electrodes with Concentric Ring Geometry for Enhancing Spatial Resolution in Electrophysiology. *Adv. Healthc. Mater.* **2017**, *6*, 1700552, doi:10.1002/adhm.201700552.
2. Zena-Giménez, V.; Garcia-Casado, J.; Ye-Lin, Y.; Garcia-Breijo, E.; Prats-Boluda, G. A flexible multiring concentric electrode for non-invasive identification of intestinal slow Waves. *Sensors* **2018**, *18*, 396.
3. Nasrollahhosseini, S.H.; Mercier, J.; Fischer, G.; Besio, W. Electrode-Electrolyte Interface Modeling and Impedance Characterizing of Tripolar Concentric Ring Electrode. *IEEE Trans. Biomed. Eng.* **2019**, *66*, 2897–2905, doi:10.1109/TBME.2019.2897935.

4. Toole, C.; Martinez-Juárez, I.E.; Gaitanis, J.N.; Sunderam, S.; Ding, L.; DiCecco, J.; Besio, W.G. Source localization of high-frequency activity in tripolar electroencephalography of patients with epilepsy. *Epilepsy Behav.* **2019**, *101*, 106519, doi:10.1016/j.yebeh.2019.106519.
5. Makeyev, O.; Musngi, M.; Moore, L.; Ye-Lin, Y.; Prats-Boluda, G.; Garcia-Casado, J. Validating the Comparison Framework for the Finite Dimensions Model of Concentric Ring Electrodes Using Human Electrocardiogram Data. *Appl. Sci.* **2019**, *9*, 4279, doi:10.3390/app9204279.
6. Garcia-Casado, J.; Ye-Lin, Y.; Prats-Boluda, G.; Makeyev, O. Evaluation of Bipolar, Tripolar, and Quadripolar Laplacian Estimates of Electrocardiogram via Concentric Ring Electrodes. *Sensors* **2019**, *19*, 3780, doi:10.3390/s19173780.
7. Aghaei-Lasboo, A.; Inoyama, K.; Fogarty, A.S.; Kuo, J.; Meador, K.J.; Walter, J.J.; Le, S.T.; Graber, K.D.; Razavi, B.; Fisher, R.S. Tripolar concentric EEG electrodes reduce noise. *Clin. Neurophysiol.* **2020**, *131*, 193–198.
8. Liu, X.; Makeyev, O.; Besio, W. Improved Spatial Resolution of Electroencephalogram Using Tripolar Concentric Ring Electrode Sensors. *J. Sens.* **2020**, *2020*, 6269394, doi:10.1155/2020/6269394.
9. Besio, W.G.; Koka, K.; Aakula, R.; Dai, W. Tri-polar concentric ring electrode development for Laplacian electroencephalography. *IEEE Trans. Biomed. Eng.* **2006**, *53*, 926–933.
10. Makeyev, O.; Ding, Q.; Besio, W.G. Improving the accuracy of Laplacian estimation with novel multipolar concentric ring electrodes. *Measurement* **2016**, *80*, 44–52, doi:10.1016/j.measurement.2015.11.017.
11. Makeyev, O.; Besio, W.G. Improving the Accuracy of Laplacian Estimation with Novel Variable Inter-Ring Distances Concentric Ring Electrodes. *Sensors* **2016**, *16*, 858, doi:10.3390/s16060858.
12. Makeyev, O. Solving the general inter-ring distances optimization problem for concentric ring electrodes to improve Laplacian estimation. *BioMed. Eng. OnLine* **2018**, *17*, 117, doi:10.1186/s12938-018-0549-6.
13. Makeyev, O.; Lee, C.; Besio, W.G. Proof of concept Laplacian estimate derived for noninvasive tripolar concentric ring electrode with incorporated radius of the central disc and the widths of the concentric rings. *Conf. Proc. IEEE Eng. Med. Biol. Soc.* **2017**, *2017*, 841–844, doi:10.1109/EMBC.2017.8036955.
14. Huiskamp, G. Difference formulas for the surface Laplacian on a triangulated surface. *J. Comput. Phys.* **1991**, *95*, 477–496, doi:10.1016/0021-9991(91)90286-T.
15. King, M.R.; Mody, N.A. *Numerical and Statistical Methods for Bioengineering: Applications in MATLAB*; Cambridge University Press: Cambridge, UK, 2010.

Publisher's Note: MDPI stays neutral with regard to jurisdictional claims in published maps and institutional affiliations.



© 2020 by the authors. Licensee MDPI, Basel, Switzerland. This article is an open access article distributed under the terms and conditions of the Creative Commons Attribution (CC BY) license (<http://creativecommons.org/licenses/by/4.0/>).